

From: [Warr \(Kettler\), Kristie](#)
To: [Moore, Gary](#)
Cc: [Cobb, Derrick](#)
Subject: CES Analytical Update
Date: Wednesday, September 10, 2014 11:11:54 AM
Attachments: [CES Sample Tracking 20140910.xlsx](#)
[tc53867 prelim X2.pdf](#)
[TC53874 complete prelim X2.PDF](#)
[TC53875 level 2.PDF](#)
[TC53876 complete prelim X2.PDF](#)

Gary,

Please see the attachments. The spreadsheet is updated with all of this information.

SDG 53875 (VB640, VB25141) – Level 2 report

SDG 53874 (VB281842, VB641, VB633) – complete prelim report, this version includes 2 sulfide results. We are waiting on the lab to correct. Once this is corrected, a level 2 report will be issued.

SDG 53876 (VB610 totals; TCLP is in 53868 that was provided yesterday) – complete prelim report, this version includes 2 sulfide results. This is being corrected. Additionally, the pesticide analyte DDT is reported as ND. This result is being verified with re-analysis. There were QC failures.

SDG 53867 (VB607 totals; TCLP is in 53868 that was provided yesterday) – partial prelim report, BTU has not been reported yet. Additionally, pesticide analyte Endosulfan I exceeded the calibration range of the instrument. The E flagged result is reported. This result will change after the dilution analysis is completed.

Please let me know if you have any questions.

Kristie

Kristie Warr

Weston Solutions, Inc.

5599 San Felipe, Suite 700

Houston, TX 77056

email: K.Warr@WestonSolutions.com

Office: 713.985.6636

Cell: 832.444.7976

CONFIDENTIALITY: This email and attachments may contain information which is confidential and proprietary. Disclosure or use of any such confidential or proprietary information without the written permission of Weston Solutions, Inc. is strictly prohibited. If you received this email in error, please notify the sender by return e-mail and delete this email from your system. Thank you.